



राजपुर गाउँपालिका
गाउँ कार्यपालिकाको कार्यालय

गंगदी , दाङ
५ न. प्रदेश, नेपाल

Technical Specification of WHO-PQS Vaccine refrigerator freezer

| S.N | CRITERIA | REQUIREMENT | OFFERED SPECIFICATION |
|-----------------------------|----------------|--|---|
| VACCINE REFRIGERATOR | | | |
| 1 | General | Refrigerant should be WHO- PQS approved having thermometers for both internal and external temperature reading | Model: WHO -PQS Code: Country of origin : |
| 2 | Capacity | Gross capacity not less than 150 litres Vaccine storage capacity not less than 110 litres | |
| 3 | Hold over time | Minimum 55 hours | |
| 4 | Refrigerant | CFC free | |
| 5 | Warranty | Minimum 2 years | |
| CHEST FREEZER | | | Model: WHO -PQS Code: Country of origin : |
| 1 | Capacity | Gross capacity not less than 275 litres Ice pack storage capacity :not less than 256 pieces of 0.6 litres | |
| 2 | Hold over time | Minimum 5 hours | |
| 3 | Refrigerant | CFC free | |
| 4 | Warranty | Minimum 2 years | |
| Voltage stabilizer | | | |
| 1 | Capacity | 1000VA 220 volt | |